

BOSE INSTITUTE  
KOLKATA

**CORRIGENDUM dt. 14.3.2014**

to Tender Notice No.  
**BI/T/11/2013-14 dt. 17.02.2014**  
*for*  
**Confocal Microscopy System**

<u>Existing specification</u>	Specifications as <b>Amended</b>
<b><u>Laser modules:</u></b>	<b><u>Laser modules:</u></b>
Combination of following single-line lasers:  1. 405 nm 2. 470 - 480 nm 3. 630 - 640 nm Each laser line must have a minimum of 5000 hr lifetime. All lasers lines are to be connected to the scan head through fibre optic cable and should be computer controlled for fast laser switching and attenuation in synchronization with the scanner.	<b><u>Combination of following laser lines should be read as :</u></b>  1. 405 nm <b>2. 470 - 490 nm</b> 3. 540 - 565 nm 4. 630 - 640 nm Each laser line must have a minimum of 5000 hr lifetime. All lasers lines are to be connected to the scan head through fibre optic cable and should be computer controlled for fast laser switching and attenuation in synchronization with the scanner.

The other terms and conditions of the tender will remain unchanged.

*Registrar, B.I.*